

Synmax Vinyl Nitrile Hybrid Medical Exam Gloves

Blue | Medical Grade | Powder Free | FDA 510(k) | Latex Free

Synmax Examination Gloves

Revolutionary new formula of vinyl-based hybrid material, containing nitrile, uses properties with elastic and strong “bones” of gloves. They offer enhanced tactile sensitivity and a practical barrier of protection for a wide range of applications. They are also a good alternative for those suffering from Type I allergies.



Features

- Brand new product
- DOP,DEHP-free
- No latex proteins, low allergy potential
- Comfortable conforming latex-like fit
- Effective barrier and extra strength
- Enhanced tactile sensitivity
- Economical protection
- Resistance to acid, alkali and oil
- 100 gloves per box

Quality standards

- Complies with **EN 455 and EN 374**
- Complies with **ASTM D5250** (USA Related Product)
- Tensile Strength: 18MPa
- Palm Thickness: 0.08 mm minimum
- Finger thickness: **0.09 mm minimum (3.5-4 mil)**
- Weight: 5.0 gm (size medium)
- Length: 230 mm minimum



Technical information

Type	Powder-free, unsterile
Material	Compound of Vinyl-based Hybrid
Colour	Blue, White, Purple, Pink
Model	Ambidextrous, with rolled rim
Storage	Protect from heat, humidity, strong light and ozone
Size/overall length as per EN 455-2	XS, S, M, L and XL 240 mm (≥230mm for USA)
Impermeability as per EN 455-1	AQL 1.5
Durability, in original package if stored as per din 7716, ISO 2230	Min. 5 years



FDA Listing for Basic Medical Synmax Vinyl Gloves

510(k) Premarket Notification

[FDA Home](#) [Medical Devices](#) [Databases](#)



[510\(k\)](#) | [DeNovo](#) | [Registration & Listing](#) | [Adverse Events](#) | [Recalls](#) | [PMA](#) | [HDE](#) | [Classification](#) | [Standards](#)
[CFR Title 21](#) | [Radiation-Emitting Products](#) | [X-Ray Assembler](#) | [Medsun Reports](#) | [CLIA](#) | [TPLC](#)

[New Search](#)

[Back To Search Results](#)

Device Classification Name	Vinyl Patient Examination Glove
510(K) Number	K153028
Device Name	Synmax Synthetic Patient Examination Vinyl Gloves, Powder Free, Blue
Applicant	ZIBO HUIYING MEDICAL PRODUCTS, CO. LTD. NO. 3 DA WU ROAD Linzi, CN 255400
Applicant Contact	Fufang Wang
Correspondent	BASIC MEDICAL INDUSTRIES, INC. 12390 East End Ave Chino, CA 91710
Correspondent Contact	Sophie Hao
Regulation Number	880.6250
Classification Product Code	LYZ
Date Received	10/16/2015
Decision Date	02/05/2016
Decision	Substantially Equivalent (SESE)
Regulation Medical Specialty	General Hospital
510k Review Panel	General Hospital
Summary	Summary
Type	Traditional
Reviewed By Third Party	No
Combination Product	No